

## Control set CLIA EBV EBNA-1 IgM

**EAN Code:** 8595635310904

**Catalog number:** CL-EBMCON

**Package size:** 2x20 tests

**Storage:** 2-8 °C

**Producer:** TestLine Clinical Diagnostics s.r.o.



- The kit is intended for use together with CLIA kits on the KleeYa® analyzer.
- The kit includes positive and negative control sera, which are used to verify the functionality of CLIA tests and the accuracy of obtained results.

For more information, please refer to [BiovendorGroup CLIA webpage](#).