



# EU DECLARATION OF CONFORMITY



Manufacturer: TestLine Clinical Diagnostics s.r.o.  
Křižíkova 188/68 CRN: 47913240  
612 00 Brno – Královo Pole  
Czech Republic, CZ-EU Tel: +420 549 121 259

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Product: Laboratory sample dryer  
Trade name: Microblot-Array dryer  
Product description: The device is used for drying of a microtiter plates.  
Power voltage 110-240 V AC, 50/60 Hz, input 240 W.

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation.

TestLine Clinical Diagnostics s.r.o. hereby declares that the respective product was tested according to:

Government Regulation No. 118/2016 Coll., which is equivalent to Directive No. 2014/35/EU of the European Parliament and of the Council, Government Regulation No. 117/2016 Coll., as amended by Government Regulation No. 350/2022 Coll., which is equivalent to Directive No. 2014/30/EU of the European Parliament and of the Council, CSN EN 61010-1 ed. 2, CSN EN 61000-6-3 ed. 2, CSN EN 55014-1 ed. 4.

The company has introduced the quality management system in accordance with the following standards:

- CSN EN ISO 9001
- CSN EN ISO 13485

The file with all the documentation is kept with the manufacturer.

## TestLine Clinical Diagnostics s.r.o.

Date and place of issue: 18.09. 2024, Brno, CZ  
Name: Bc. Lenka Hanáková  
Position: Quality Manager



TestLine Clinical Diagnostics s.r.o.  
QUALITY CONTROL  
Krizikova 68, 612 00 BRNO  
tel.: +420549121219, tel./fax: +420541243390  
VAT: CZ47913240  
Czech Republic