



EU DECLARATION OF CONFORMITY



Manufacturer: TestLine Clinical Diagnostics s.r.o.
Křižíkova 188/68 CRN: 47913240
612 00 Brno – Královo Pole
Česká republika CZ-EU Tel: +420 549 121 259

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Product: Laboratory sample dryer
Type designation: SV-190-170-134
Trade name: Microblot-Array dryer
Product description: The device is used for drying of a microtiter plates.
Power voltage 230 V AC, 50 Hz, input 240 W.

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation.

TestLine Clinical Diagnostics s.r.o. hereby declares that the respective product was tested according to:

Government regulation no. 118/2016 Coll., which is equivalent to directive no. 2014/35/EU of the European Parliament and of the Council, Government regulation no. 117/2016 Coll., which is equivalent to directive no. 2014/30/EU of the European Parliament and of the Council, ČSN EN 61010-1 ed. 2, ČSN EN 61000-6-3 ed. 2, ČSN EN 55014-1 ed. 4.

The company has introduced the QM system in accordance with the following standards:

- ČSN EN ISO 9001
- ČSN EN ISO 13485

The file with all the documentation is kept with the manufacturer.

TestLine Clinical Diagnostics s.r.o.

Date and place of issue: 16.08. 2021, Brno, CZ
Name: Bc. Lenka Hanáková
Position: Quality Manager



TestLine Clinical Diagnostics s.r.o.
QUALITY CONTROL
Krizikova 68, 612 00 BRNO
tel.: +420549121219, tel./fax: +420541243390
VAT: CZ47913240
Czech Republic