

## DS2



**Catalog number:** 62010  
**Storage:** Bez teplotního režimu  
**Producer:** Dynex Technologies, Inc.

### Samples

- Capacity: 100 serum tubes; diameter: 10-16 mm; height: 40-100 mm

### Robotic and pipetting system

- 1 pipetting arm for replaceable tips (2 types)
- Processing capacity: 2 plates
- Pipetting volume: 10-250  $\mu$ l, 20-1000  $\mu$ l
- Volume control system

### Washing

- 1 station
- 2 wash solutions with liquid-level monitoring
- Waste: integrated container (1,5 l)
- Number of channels: 8-channel wash head
- Dispensing volume range: 50-1000  $\mu$ l
- Time of fluid in well: 0-999 s
- Residual volume: < 0,3  $\mu$ l

## Measurement

- 1 photometer, 0-3,0 OD, 405-690 nm
- 6 interference filters
- Number of optical channels: 12

## Incubation

- 2 independent incubation units
- Temperature range: RT; 4 °C – 40 °C
- Stability:  $\pm 1$  °C
- Possibility of shaking during incubation (different speed setting)

## Software

- A part of PC delivery, OS Windows 7
- Archiving patients results
- Quality control by Levey-Jennings, Westgard rules

## Safety

- Liquid-level monitoring
- Bubbles and clots detection
- Barcode reader – errors elimination
- Acoustic notifications

## Dimensions

- Depth: 68 cm; height: 66 cm; width: 54 cm
- Weight: 48 kg

## Certification

- CE marked, IVD certified