

# EIA EBV EA-D IgM



EAN Code: 8595635302718 Catalog number: EAM096 Package size: 96 tests

Storage: 2-8 °C

**Producer:** TestLine Clinical Diagnostics s.r.o.

## **Description:**

- Microtitre wells are coated with the purified and inactivated EBV EA-D antigen.
- If specific antibodies are present, they bind to the antigen, are labeled by the Conjugate in the following steps and are detected by color reaction with a single component substrate (TMB-Complete).
- The kit allows 96 tests, including controls in a split microtiter plate with color-coded strips and breakable wells.

#### Advantages:

- The total assay time is about 3 hours.
- High sensitivity and specificity of the test.
- Kit includes CUT-OFF, Positive Control and Negative Control, Calibrators.
- Semi-quantitative evaluation in the Index of Positivity (IP).
- Ready-to-use, color-coded components.
- Single-component substrate.
- Interchangeable components with the exception of kit specific components (Controls, Conjugate,

Exportováno 31. 3. 2025 Stránka 1 / 2



Plate).

• Sample diluent contains IgG/RF sorbent.

## **Application:**

- Screening tests for the detection of infection with EBV and determination of infectious stage in humans.
- Disease phase diagnostics.

### Brief assay procedure:

- 1. Dilute samples (1:101).
- 2. Pipette Controls and diluted Samples.
- 3. Incubate at 37°C for 60 min.
- 4. Aspirate and wash the wells 5×
- 5. Pipette Conjugate.
- 6. Incubate at 37°C for 60 min.
- 7. Aspirate and wash the wells 5×
- 8. Pipette Substrate (TMB-Complete).
- 9. Incubate at 37°C for 30 min.
- 10. Pipette Stop Solution.
- 11. Read color intensity at 450 nm
- 12. Evaluate the test.

Exportováno 31. 3. 2025 Stránka 2 / 2