

## Microblot-Array Yersinia IgG



EAN Code: 8595635310683 Catalog number: YGMA048 Package size: 48 tests

Storage: 2-8 °C

**Producer:** TestLine Clinical Diagnostics s.r.o.

## **Description:**

- Used Antigen: combination of selected parts of the specific antigens of Yersinia enterocolitica (YopB, YopD, YopM, YopN, LcrV, Ail, Invasin, YscM-Y.Ent).
- Recombinant antigens are applied in the form of microdots (spots) in triplets to a nitrocellulose membrane in a microtiter well format.
- If specific antibodies are present in the sample, they will bind to the respective antigens.
- The complex is labelled with Conjugate and detected through a colour reaction with substrate (BCIP/NBT).
- The kit enables 48 tests.
- Short incubation periods, total assay time: approximately 1.5 h.
- Each individual test includes control spots, which serves to check functionality and sensitivity of the sets, and also the spots for checking the presence of the conjugate.
- Positive controls included enabling validation of the tests.
- For quantitative evaluation the wells contain calibration spots.
- All reagents supplied are ready to use.
- The kit may be used sequentially for smaller batches of samples (reagents are provided in sufficient quantities).
- Evaluation using Microblot-Array reader and the appropriate software.

## **Applications:**

- Diagnostics of Yersinia infection.
- Confirmatory method for ELISA.

## **Brief assay procedure:**

1. Dilute samples serum/plasma (1:51).

Exportováno 29. 3. 2025 Stránka 1 / 2





- 2. Pipette diluted samples.
- 3. Incubate at room temperature for 30 min.
- 4. Wash 3 times for 5 min.
- 5. Add Conjugate.
- 6. Incubate at room temperature for 30 min.
- 7. Wash 3 times for 5 min.
- 8. Add Substrate (BCIP/NPT).
- 9. Incubate at room temperature for 15 min.
- 10. Wash 2 times for 5 min in distilled water.
- 11. Dry and evaluate of test.

Exportováno 29. 3. 2025 Stránka 2 / 2