

# **SmartEIA ASCA IgG**



**EAN Code:** 8595635306587 **Catalog number:** SK-ScG096

Package size: 96 tests

Storage: 2-8 °C

**Producer:** TestLine Clinical Diagnostics s.r.o.

#### **Description:**

- Microtitre wells are coated with high purified antigen mannan isolated from Saccharomyces cerevisiae
- If specific antibodies are present, they bind to the antigen, are labeled by the Conjugate in the following steps and are detected by color reaction with a single component substrate (TMB-Complete).
- The kit allows 96 tests, including controls in a split microtiter plate with color-coded strips and breakable wells.

# Advantages:

- The total assay time is about 2 hours.
- High sensitivity and specificity of the test.
- Kit includes CUT-OFF, Positive Control, Negative Control and Calibrators (5, 20, 80, 320 AU/ml).
- Semi-quantitative evaluation in the Index of Positivity (IP) or quantitative evaluation in AU/ml.
- Ready-to-use, color-coded components.
- Single-component substrate.
- Interchangeable components with the exception of kit specific components (Controls, Conjugate, Plate).

### **Application:**

Laboratory test for differential diagnostics of inflammatory bowel disease.

## **Brief assay procedure:**

Exportováno 29. 3. 2025 Stránka 1 / 2



- 1. Dilute samples (1:51).
- 2. Incubation at 37 °C for 30 min.
- 3. Pipette Controls and diluted Samples.
- 4. Incubate at 37°C for 15 min.
- 5. Aspirate and wash the wells 5×
- 6. Pipette Conjugate.
- 7. Incubate at 37°C for 15 min.
- 8. Aspirate and wash the wells 5×
- 9. Pipette Substrate (TMB-Complete).
- 10. Incubate at 37°C for 30 min.
- 11. Pipette Stop Solution.
- 12. Read color intensity at 450 nm.
- 13. Evaluate the test.

SmartEIA kits are specifically designed for automated processing on the Agility® instrument, Dynex Technologies, Inc.

Exportováno 29. 3. 2025 Stránka 2 / 2